

Avery Weigh-Tronix

Calibration Certificate

Avery Weigh-Tronix
Foundry Lane,
Smethwick, West Midlands
B66 2LP England
www.averyweigh-tronix.com
Tel: 0845 900 22 44

Avery Weigh-Tronix
Unit 52, Mallusk Enterprise Pk,
2 Mallusk Drive,
Newtownabbey BT36 4GN
www.averyweigh-tronix.com
Tel: 028 9083 7115

Avery Weigh-Tronix
Unit L11,
Greenogue Business Park
Rathcoole, Co. Dublin
www.averyweigh-tronix.com
Tel: 01 4000 700

Customer Details

Name: Kartworld
Address: Meenane
Watergrasshill
Co. Cork
Eire
T56 TK73
Contact: David Walsh
Telephone No. 087-2559100

Certificate No: KF020425-700

Date of Calibration: 02/04/2025
Next Calibration Due: April 2026

Agreement No:

Location:

Department:

Plant Number:

Manufacturer: Tru-Test

Model: Model 700

Serial No: 700

Capacity: 250kg

Resolution: 0.5kg

Tolerance: As Below

Calibration Frequency: 12 Months

Linearity / Accuracy Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0kg	0.0kg	Nil	0.0kg	Nil
50kg	50.0kg	Nil	50.0kg	Nil
100kg	100.5kg	+0.5kg	100.5kg	0.5kg
150kg	150.5kg	+0.5kg	150.5kg	0.5kg
200kg	201.0kg	+1.0kg	201.0kg	1.0kg
250kg	251.0kg	+1.0kg	251.0kg	1.0kg

Off Centre Loading Test 2 3 or End-Middle-End 1 4 1-2-3

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
50kg	50.5kg	+0.5kg	50.5kg	+0.5kg
50kg	50.0kg	Nil	50.0kg	Nil
50kg	50.0kg	Nil	50.0kg	Nil
50kg	49.5kg	-0.5kg	49.5kg	-0.5kg

Test Weight Details

Weight Set No. Certificate No.

T165 UK240300

Test weights and equipment are traceable to national standards

Repeatability Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
150kg	150.5kg	+0.5kg	150.5kg	0.5kg
150kg	150.5kg	+0.5kg	150.5kg	0.5kg
150kg	150.5kg	+0.5kg	150.5kg	0.5kg

Calibration is carried out to UKWF Calibration Code of Practice requirements & Avery Weigh-Tronix procedures for non-automatic weighing instruments.

Comments:

Tolerance +/- 3 Divisions.


Tolerances

Unless otherwise agreed tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:

0-500 divisions +/- 1 divn.
>500-2000 divisions +/- 2 divn.
>2000 divisions +/- 3 divn.

Technician Name: K Flanagan

Technician Signature:



Registered in the UK: ITW Limited trading as Avery Weigh-Tronix. Registered Office: Saxon House, 2-4 Victoria Street, Windsor, SL4 1EN England. No. 559693.

Registered in Ireland 22058: Avery Weigh-Tronix is a trading name of Berkel (Ireland) Ltd. Registered address: Wilton Secretarial Ltd, 1st Floor, Fitzwilton Hse, Wilton Place, Dublin 2